

---

**CERTIFICATE OF ANALYSIS****Product Information**

---

|                   |  |
|-------------------|--|
| Product name      | Human Genomic DNA  |
| Cat. No.          | M00094   |
| Lot No.           | C20051806  |
| Storage condition | Store at 4°C for frequent use. Store at -20°C for occasional use.<br>Dot not freeze-thaw more than 3 times. Avoid agitation. |

---

**QC Results**

---

| Test Items                      | Specifications                             | Results |
|---------------------------------|--|---------|
| Absorbance ration               | A260/280 $\geq$ 1.8<br>A260/230 $\geq$ 2.0 | PASS    |
| Concentration                   | $\geq$ 100 ng/ $\mu$ l                     | PASS    |
| Assaying for<br>Quality of Gels | The bands are clear                        | PASS    |

---

**Caution:**

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:  Date: 06/05/2018

**Product Department**