

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE8941603

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



I Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.334	10198	1428	0.487
2	15.579	5929	1237	0.283
3	16.033	1773	262	0.085
4	16.722	10096	1843	0.482
5	16.906	10936	1671	0.522
6	17.047	9374	1664	0.447
7	17.253	2031060	318853	96.903
8	17.762	2605	495	0.124
9	17.889	9288	1780	0.443
10	18.038	4715	968	0.225
Total		2095972	330200	100.000