

CERTIFICATE OF ANALYSIS

Product Information

Product Name:	QuickClean II PCR Extraction Kit
Cat. No.:	L00419
Lot No.:	C51021603
Exp. Date:	25 Sep 2017
Storage Condition:	RT

Kit Contents

Kit Components	50 Assays
Binding Buffer	10ml
Wash Buffer	15ml
Elution Buffer	10ml
Spin columns	50(Column Volume:750μl)

Quality Control

Test Items	Specifications	Results
A260/A280	1.8-2.0	Pass
Electrophoresis	The brightness of the 1000bp PCR band after purified more than 70% that of the band before purified	Pass

Caution

For laboratory use only. Not for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Certified by:  Date: 03/25/2016
Product Department