

Sample Name :Bradykinin  
Cat. No. :RP11129  
Lot. No. :PE6421605

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
Total Flow:1 ml/min  
Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.762	5095	795	0.079
2	15.847	4605	592	0.071
3	16.236	6376606	275562	98.475
4	18.017	32377	1296	0.500
5	20.472	4865	657	0.075
6	21.369	7102	546	0.110
7	22.389	41391	3884	0.639
8	22.790	3343	442	0.052
Total		6475385	283774	100.000