

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE1781601

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	16.598	26641	2342	0.555
2	17.034	11447	2249	0.239
3	18.220	35179	5585	0.733
4	18.406	15743	3283	0.328
5	18.555	4643619	546734	96.819
6	19.393	45786	5813	0.955
7	19.617	17786	3093	0.371
Total		4796202	569099	100.000