

Sample Name :Adrenocorticotrophic Hormone (ACTH) (1-39), rat
 Cat. No. : RP11305
 Lot. No. : E9592519L
 Time Processed : 7:54:47
 Month-Day-Year Processed : 12/19/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0309

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.663	1323	137	0.014
2	11.987	1561	222	0.017
3	12.209	6337	855	0.069
4	12.729	1367	163	0.015
5	15.681	9061258	1214229	98.608
6	15.950	36304	3634	0.395
7	16.222	43419	4483	0.473
8	17.260	2943	378	0.032
9	17.430	3569	569	0.039
10	17.656	1185	176	0.013
11	17.784	1257	220	0.014
12	17.932	11452	1907	0.125
13	21.500	2159	263	0.023
14	21.738	1505	194	0.016
15	22.458	8739	1331	0.095
16	23.608	1243	144	0.014
17	23.844	2418	202	0.026
18	24.177	1103	79	0.012
Total		9189142	1229187	100.000