

Sample Name :NP (396-404)  
 Cat No. :RP20090  
 Lot No. :PE5791903  
 Time Processed :22:36:53  
 Month-Day-Year Processed :03/29/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ17010508

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.142	7292	697	0.078
2	14.360	9719	1670	0.104
3	14.487	23053	2969	0.247
4	15.001	9050233	1256295	96.847
5	15.251	216337	22603	2.315
6	15.703	38245	6335	0.409
Total		9344880	1290569	100.000