

Sample Information

Sample Name :Substance P (7-11)
 Cat. No :RP10182
 Lot. No. :PE9921806

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>
 Flow : 1.0000 mL/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>
 <Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Peak#	Ret. Time	Height	Area	Height%	Area%
1	15.133	4644	39631	0.920	1.026
2	17.265	488968	3757254	96.843	97.309
3	17.664	609	4379	0.121	0.113
4	17.934	10689	59880	2.117	1.551
Total		504910	3861145	100.000	100.000