

Sample Name:MOG(35-55)  
 Cat No :RP10245  
 Lot No :E9382221D  
 Time Processed :5:51:58  
 Month-Day-Year Processed :04/30/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: SS-CM-0309

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.406	4103	577	0.019
2	11.510	2560	319	0.012
3	11.738	8894	692	0.041
4	12.304	76955	5693	0.357
5	12.668	448201	37498	2.079
6	12.960	20537091	2169565	95.253
7	13.213	300038	26523	1.392
8	13.951	3771	423	0.017
9	14.306	8547	1138	0.040
10	15.303	4926	529	0.023
11	15.750	16157	1480	0.075
12	16.543	2236	119	0.010
13	16.939	1758	167	0.008
14	17.602	2744	214	0.013
15	18.448	69663	3833	0.323
16	19.375	1021	164	0.005

Peak#	Ret. Time	Area	Height	Area%
17	19.501	3447	292	0.016
18	20.025	64177	3405	0.298
19	22.016	1472	145	0.007
20	23.134	1787	230	0.008
21	24.563	1065	103	0.005
Total		21560612	2253110	100.000