

Sample Name : Exendin (9-39)  
 Cat No.: RP10872  
 Lot No.: E5647520C  
 Time Processed : 16:58:08  
 Month-Day-Year Processed : 03/28/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	15
25.00	Pumps	Pump A B.Conc	75
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	15
40.00	Pumps	Pump A B.Conc	15
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	17.275	19687	3242	0.328
2	18.541	5866682	865427	97.650
3	18.996	91399	5780	1.521
4	20.056	7805	1858	0.130
5	20.807	22307	3153	0.371
Total		6007880	879460	100.000