

Sample Name : (Arg)9
 Cat No. : RP20403
 Lot No. : E8669621F
 Time Processed : 11:51:47 PM
 Month-Day-Year Processed : 06/24/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm
 <<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|---------------|-------|
| 0.01 | Pumps | Pump A B.Conc | 0 |
| 5.00 | Pumps | Pump A B.Conc | 0 |
| 20.00 | Pumps | Pump A B.Conc | 30 |
| 25.00 | Pumps | Pump A B.Conc | 65 |
| 25.01 | Pumps | Pump A B.Conc | 95 |
| 27.00 | Pumps | Pump A B.Conc | 95 |
| 27.01 | Pumps | Pump A B.Conc | 0 |
| 35.00 | Pumps | Pump A B.Conc | 0 |
| 35.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|---------|--------|---------|
| 1 | 16.038 | 13027 | 1792 | 0.365 |
| 2 | 16.458 | 3473030 | 385192 | 97.387 |
| 3 | 18.309 | 19963 | 3687 | 0.560 |
| 4 | 20.938 | 60197 | 10336 | 1.688 |
| Total | | 3566217 | 401006 | 100.000 |