

Sample Name :Abl Cytosolic Substrate  
 Cat No. :RP19985  
 Lot No. :E4830719I  
 Time Processed :23:02:36  
 Month-Day-Year Processed :09/28/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.146	14771	2661	0.333
2	12.641	4249903	717138	95.852
3	13.671	40035	6676	0.903
4	16.428	125698	21453	2.835
5	20.484	3426	405	0.077
Total		4433832	748333	100.000