


Mass Spectrum



Sample Information

User : Gary
Month-Day Processed : 04/01/20
Time Processed : 06:41:56 PM
Injection Volume : 0.3
Sample Name : MUC5AC
Cat No. : RP20290
Lot No. : E8625820D
Theoretical MW : 1501.62
Observed MW : 1501.2

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200

Equipment : GK11010007
Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H₂O/50%MeOH