

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE8531702

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : UniPS 10-300 10µm 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	16.086	4232	899	0.246
2	16.415	24387	3408	1.420
3	17.058	9996	1173	0.582
4	17.302	4494	1024	0.262
5	17.515	1649594	312562	96.030
6	17.985	3902	1149	0.227
7	19.580	9388	1104	0.547
8	20.345	11800	1442	0.687
Total		1717793	322761	100.000