

Sample Name : Bradykinin (2-9)  
 Cat. No. : RP11135  
 Lot. No. : E4786819I  
 Time Processed : 14:38:18  
 Month-Day-Year Processed : 09/21/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	0	
5.00	Pumps	Pump A B.Conc	0	
20.00	Pumps	Pump A B.Conc	30	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	0	
40.00	Pumps	Pump A B.Conc	0	
40.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GK11010011

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	16.030	131172	15678	2.106
2	19.965	41211	4475	0.662
3	22.215	21242	2901	0.341
4	23.658	8769	2035	0.141
5	23.854	6016223	628868	96.580
6	24.242	10634	1178	0.171
Total		6229252	655134	100.000