

Sample Name :NP (396-404)
 Cat. No. : RP20090
 Lot. No. : E6466220A
 Time Processed :8:03:39
 Month-Day-Year Processed : 01/12/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.00	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0286



1 Det.A Ch1 / 220nm
 Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.735	1147	123	0.025
2	14.069	4576	771	0.098
3	14.219	9447	1521	0.202
4	14.448	4751	971	0.102
5	14.672	10180	1760	0.218
6	14.894	4490384	688799	95.961
7	15.146	87866	14307	1.878
8	15.217	51740	9088	1.106
9	15.549	16342	2737	0.349
10	17.736	1773	108	0.038
11	23.647	1171	156	0.025
Total		4679377	720342	100.000