

Sample Name :Exendin (9-39)  
 Cat No. :RP10872  
 Lot No. :E0673721A  
 Time Processed :10:55:49  
 Month-Day-Year Processed :02/06/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	17.361	78578	10646	0.919
2	18.749	15516	1905	0.182
3	21.921	120960	7634	1.415
4	22.596	90397	7008	1.058
5	22.901	8167951	526772	95.556
6	23.637	74400	5455	0.870
Total		8547803	559419	100.000