

Sample Name :Magainin 2  
 Cat. No. : RP11232  
 Lot. No. : PE8131803

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	16.366	11163	1005	0.720
2	16.878	2915	356	0.188
3	17.305	1486248	169369	95.897
4	17.618	8534	1569	0.551
5	17.841	21297	2122	1.374
6	18.000	7389	1013	0.477
7	18.560	3985	695	0.257
8	18.698	4366	764	0.282
9	20.457	3938	526	0.254
Total		1549835	177419	100.000