

Sample Name :InsB 9-23
 Cat No. :RP20391
 Lot No. :E39866201
 Time Processed :5:25:36
 Month-Day-Year Processed :09/19/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.884	29088	5402	0.866
2	16.719	17335	2732	0.516
3	17.044	11622	1388	0.346
4	17.428	3197788	443110	95.216
5	18.486	34294	3886	1.021
6	19.116	68328	9821	2.035
Total		3358456	466338	100.000