

Sample Name : Leuprolide Acetate
 Cat. No. : RP17321
 Lot. No. : PE8631902
 Time Processed : 9:22:53
 Month-Day-Year Processed : 03/23/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0309

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.649	7253	570	0.077
2	13.823	11352	631	0.120
3	14.892	209024	41711	2.214
4	15.984	6347	1226	0.067
5	16.534	29921	4993	0.317
6	16.658	9877	1281	0.105
7	16.977	24482	4207	0.259
8	17.269	72260	11423	0.765
9	17.448	9042157	1537947	95.780
10	17.728	27835	5704	0.295
Total		9440507	1609693	100.000