

Sample Name :Magainin 2
 Cat No. :RP11232
 Lot No. :E9622419L
 Time Processed :7:50:10
 Month-Day-Year Processed :12/19/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	

<<Column Performance>>

<Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ17010507

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.799	29056	5522	0.514
2	15.926	22725	4665	0.402
3	16.847	47182	9578	0.835
4	17.041	5368653	839863	95.058
5	17.262	177151	23666	3.137
6	24.921	2985	498	0.053
Total		5647751	883791	100.000