

Sample Name : Amylin (8-37), amide, rat

Cat. No. : RP11277 Lot. No. : PE9881806

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow: 1 ml/min

Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
10.00	Pumps	Pump A B.Conc	35	
35.00	Pumps	Pump A B.Conc	95	
41.00	Pumps	Pump A B.Conc	95	
41.01	Pumps	Pump A B.Conc	5	
50.00	Pumps	Pump A B.Conc	5	
50.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	6.121	154851	18130	3.133
2	13.922	3318	434	0.067
3	14.086	18698	1397	0.378
4	14.692	3660	511	0.074
5	15.017	32000	3271	0.647
6	15.213	4702127	494841	95.122
7	15.483	12438	2984	0.252
8	16.142	16166	1239	0.327
Total		4943260	522809	100.000