

Sample Name :LL-37, scrambled
 Cat No. :RP20332
 Lot No. :E6891020H
 Time Processed :7:44:53
 Month-Day-Year Processed :08/06/20

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	5.404	10551	1488	0.802
2	15.249	1277103	113800	97.116
3	15.768	2833	482	0.215
4	16.489	22382	3316	1.702
5	20.520	1084	130	0.082
6	22.030	1077	122	0.082
Total		1315031	119338	100.000