


Mass Spectrum



Sample Information

User : Gary
Month-Day Processed : 09/30/19
Time Processed : 12:49:11 PM
Injection Volume : 0.3
Sample Name : β -Amyloid (1-9)
Cat No. : RP20160
Lot No. : E4841619I
Theoretical MW : 1033.00
Observed MW : 1032.8

Equipment : GK11010007
Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H₂O/50%MeOH

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250
Block Temp : 200