

Sample Name : Bradykinin (1-7)
 Cat. No. : RP11133
 Lot. No. : E7304419L
 Time Processed :1:07:22
 Month-Day-Year Processed :12/12/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GR11010440

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.204	6229	956	0.124
2	12.342	6320	586	0.126
3	12.511	19183	2228	0.383
4	12.888	4971314	518112	99.171
5	17.731	7481	1014	0.149
6	24.836	2354	291	0.047
Total		5012882	523186	100.000