

Sample Name : [Leu⁵]-Enkephalin
 Cat. No. : RP10884
 Lot. No. : PE8211701

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.328	13508	2864	0.059
2	16.464	815	174	0.004
3	16.792	8109	1145	0.036
4	16.922	25790	4928	0.113
5	17.090	22646431	2012312	99.227
6	17.417	44909	12426	0.197
7	17.507	40510	8273	0.178
8	24.289	42691	5621	0.187
Total		22822763	2047743	100.000