

Sample Name :LL-37, scrambled  
 Cat No :RP20332  
 Lot No :E8667120L  
 Time Processed :8:52:27 PM  
 Month-Day-Year Processed :12/05/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.00	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.882	249572	54411	1.598
2	14.483	158658	19061	1.016
3	14.954	146054	31502	0.935
4	15.351	14843957	2900415	95.048
5	15.716	38831	6964	0.249
6	17.059	94308	17771	0.604
7	18.140	85933	13982	0.550
Total		15617314	3044106	100.000