

Sample Name :Magainin 2
 Cat No. :RP11232
 Lot No. :E46525201
 Time Processed :14:40:42
 Month-Day-Year Processed :09/21/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.247	24354	2485	0.591
2	15.345	28202	2664	0.684
3	16.287	3916514	611913	95.010
4	16.524	99491	13693	2.414
5	16.973	38101	6100	0.924
6	17.148	15532	2156	0.377
Total		4122194	639011	100.000