

Sample Name :InsB 9-23
 Cat No. :RP20391
 Lot No. :E4697320K
 Time Processed :3:07:30 AM
 Month-Day-Year Processed :11/13/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
41.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.992	2413	333	0.027
2	14.373	8903963	1184139	99.183
3	14.629	22736	6457	0.253
4	15.209	16683	1284	0.186
5	15.481	12117	1501	0.135
6	15.666	19351	2651	0.216
Total		8977263	1196364	100.000