

Sample Name: NP (396-404)
 Cat. No. : RP20090
 Lot. No. : E7107821D
 Time Processed: 1:37:57
 Month-Day-Year Processed :04/04/2021
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm
 <<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment:ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.660	1577	198	0.020
2	13.068	1925	288	0.024
3	13.282	3197	403	0.041
4	13.418	3378	454	0.043
5	13.623	11313	1131	0.144
6	14.041	19955	2161	0.254
7	14.400	26349	2868	0.335
8	14.606	7701792	814196	97.913
9	15.301	2554	346	0.032
10	15.887	28517	3782	0.363
11	16.505	1353	141	0.017
12	18.425	1311	164	0.017
13	18.728	27102	1762	0.345
14	20.951	4943	348	0.063
15	21.665	2620	225	0.033
16	23.651	24006	916	0.305

Peak#	Ret. Time	Area	Height	Area%
17	24.574	2256	193	0.029
18	24.836	1788	209	0.023
Total		7865935	829785	100.000