

Sample Name :GRGDSP  
 Cat. No. : RP20255  
 Lot. No. : E4192519F  
 Time Processed : 18:56:38  
 Month-Day-Year Processed :06/11/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ17010507

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	4.393	72147	8448	0.676
2	4.991	2008	357	0.019
3	5.302	16911	2091	0.158
4	5.634	10552045	532787	98.890
5	7.933	1511	262	0.014
6	8.434	2475	516	0.023
7	8.604	9724	1940	0.091
8	10.313	3032	516	0.028
9	13.363	8867	309	0.083
10	14.124	1802	175	0.017
Total		10670520	547402	100.000