

Sample Name :RGD
 Cat. No. : RP20297
 Lot. No. : PE5111706


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.234	39442	5177	1.404
2	10.781	2687592	469180	95.672
3	11.085	11679	2797	0.416
4	11.303	6707	1116	0.239
5	11.512	13712	1981	0.488
6	12.193	34165	6330	1.216
7	19.734	15887	2658	0.566
Total		2809185	489239	100.000