

Sample Name : CEF6
 Cat No.: RP11768
 Lot No.: E4027021F
 Time Processed : 13:37:45
 Month-Day-Year Processed : 06/16/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.833	26732	3301	0.279
2	12.705	32718	4470	0.341
3	13.339	14618	2123	0.153
4	14.011	9350413	806845	97.579
5	14.338	149451	26728	1.560
6	15.291	8503	1551	0.089
Total		9582435	845019	100.000