

Sample Name :Magainin 2  
 Cat No :RP11232  
 Lot No :E5073720J  
 Time Processed :1:17:17 AM  
 Month-Day-Year Processed :10/14/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm


| Time  | Module     | Command | Value |
|-------|------------|---------|-------|
| 0.01  | Pumps      | B.Conc  | 5     |
| 25.00 | Pumps      | B.Conc  | 65    |
| 25.01 | Pumps      | B.Conc  | 95    |
| 31.00 | Pumps      | B.Conc  | 95    |
| 31.01 | Pumps      | B.Conc  | 5     |
| 40.00 | Pumps      | B.Conc  | 5     |
| 40.01 | Controller | Stop    |       |

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area    | Height | Area %  |
|-------|-----------|---------|--------|---------|
| 1     | 17.649    | 203001  | 15386  | 2.616   |
| 2     | 18.066    | 7558281 | 768385 | 97.384  |
| Total |           | 7761282 | 783770 | 100.000 |