


Sample Name: Substance P (7-11)
 Cat No. : RP10182
 Lot No. : PE5191707
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	17.092	2455	327	0.037
2	17.295	27168	2194	0.412
3	17.637	2836	352	0.043
4	17.906	56351	5284	0.854
5	18.136	30648	3891	0.465
6	18.380	6340462	439924	96.095
7	19.176	92170	5822	1.397
8	19.574	3910	343	0.059
9	20.675	42117	3869	0.638
Total		6598118	462006	100.000