

Sample Name :MOG(35-55)
 Cat. No. :RP10245
 Lot. No. :E7291319H
 Time Processed :19:20:42
 Month-Day-Year Processed :09/01/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.900	33803	4386	0.512
2	14.333	3193	618	0.048
3	14.482	22571	2894	0.342
4	14.682	189487	18777	2.870
5	15.045	6346921	974917	96.131
6	20.373	2459	387	0.037
7	21.547	3930	562	0.060
Total		6602364	1002541	100.000