

Sample Name :NP (396-404)
 Sample ID :U5187DD070-1

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	10.345	3261	592	0.221
2	10.623	5002	746	0.339
3	11.317	23546	3773	1.596
4	14.995	1698	219	0.115
5	15.217	10554	800	0.715
6	15.531	1402332	202721	95.052
7	15.806	26637	2526	1.806
8	21.038	2303	284	0.156
Total		1475333	211661	100.000