

Sample Name : PAR-4 Agonist Peptide, amide  
 Cat. No. : RP11529  
 Lot. No. : PE2581806


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.582	15718	1639	0.181
2	8.959	33146	3013	0.382
3	9.629	9244	1043	0.107
4	10.092	22289	1886	0.257
5	12.392	9142	1245	0.105
6	12.614	8507483	997275	98.143
7	12.883	42121	5338	0.486
8	13.545	29335	3956	0.338
Total		8668478	1015396	100.000