

Sample Name : CEF6
 Cat No. : RP11768
 Lot No. : E4725619I
 Time Processed : 17:55:24
 Month-Day-Year Processed : 09/19/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.226	3602	354	0.043
2	11.369	7577	572	0.090
3	12.180	249517	11586	2.953
4	12.904	14244	1591	0.169
5	13.082	22116	1933	0.262
6	13.306	46559	3610	0.551
7	13.895	8068626	708660	95.488
8	14.477	9968	1816	0.118
9	14.779	16565	2169	0.196
10	15.050	11074	1514	0.131
Total		8449848	733805	100.000