

Sample Name : Leuprolide Acetate  
 Cat. No. : RP17321  
 Lot. No. : E0771819L  
 Time Processed : 7:50:42  
 Month-Day-Year Processed : 12/12/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
27.00	Pumps	Pump A B.Conc	95	
27.01	Pumps	Pump A B.Conc	5	
35.00	Pumps	Pump A B.Conc	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GK11010011

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.291	13039	2149	0.134
2	14.339	1861	271	0.019
3	16.100	60156	5843	0.618
4	16.792	51211	4165	0.526
5	16.955	33166	3893	0.341
6	17.273	9538716	1352495	97.975
7	17.608	26638	2948	0.274
8	18.356	11122	1271	0.114
Total		9735908	1373035	100.000