

Sample Name :LL-37, scrambled
 Cat No. :RP20332
 Lot No. :E5114320F
 Time Processed :7:24:44
 Month-Day-Year Processed :06/05/20

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|-----------------|-------|
| 0.01 | Pumps | Solvent B Conc. | 5 |
| 25.00 | Pumps | Solvent B Conc. | 65 |
| 25.01 | Pumps | Solvent B Conc. | 95 |
| 27.00 | Pumps | Solvent B Conc. | 95 |
| 27.01 | Pumps | Solvent B Conc. | 5 |
| 35.00 | Pumps | Solvent B Conc. | 5 |
| 35.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|---------|--------|---------|
| 1 | 13.614 | 12018 | 268 | 0.527 |
| 2 | 15.336 | 2262804 | 195110 | 99.142 |
| 3 | 15.857 | 2868 | 517 | 0.126 |
| 4 | 20.552 | 2774 | 204 | 0.122 |
| 5 | 22.858 | 1924 | 200 | 0.084 |
| Total | | 2282388 | 196298 | 100.000 |