

Sample Name :Bradykinin (1-3)
 Cat No. :RP11130
 Lot No. :E7489921B
 Time Processed : 14:22:15
 Month-Day-Year Processed : 02/24/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	B.Conc	15
25.00	Pumps	B.Conc	75
25.01	Pumps	B.Conc	95
32.00	Pumps	B.Conc	95
32.01	Pumps	B.Conc	15
40.00	Pumps	B.Conc	15
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: GK11010017

mV



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	8.837	24902	3386	0.114
2	9.107	551688	65833	2.529
3	9.516	21117668	3075565	96.819
4	9.779	117343	14632	0.538
Total		21811601	3159417	100.000