

Sample Name:Neuropeptide FF Morphine Modulating Neuropeptide
 Cat. No. : RP10470
 Lot. No. : E6457219G
 Time Processed :1:30:08 PM
 Month-Day-Year Processed :07/25/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.049	16721	532	0.203
2	10.004	1583	219	0.019
3	11.206	5214	801	0.063
4	12.373	1153	107	0.014
5	14.026	8117525	1151268	98.515
6	14.419	32479	6494	0.394
7	15.879	53360	5319	0.648
8	18.079	2795	273	0.034
9	19.150	1313	80	0.016
10	21.030	1965	255	0.024
11	22.906	1032	87	0.013
12	23.617	1268	97	0.015
13	23.928	1304	90	0.016
14	24.644	2216	114	0.027
Total		8239929	1165734	100.000