

Sample Name: MOG(35-55)  
 Cat No. : RP10245  
 Lot No. : E3497619K  
 Time Processed : 23:02:55  
 Month-Day-Year Processed : 11/23/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.210	86979	13940	0.652
2	14.215	6171	741	0.046
3	14.333	2120	446	0.016
4	14.837	325197	39018	2.439
5	15.155	12896926	1463560	96.717
6	16.090	3829	552	0.029
7	16.365	3189	379	0.024
8	19.129	1359	72	0.010
9	20.995	1433	135	0.011
10	23.703	3039	398	0.023
11	24.838	4487	610	0.034
Total		13334731	1519851	100.000