

Sample Name :Cecropin B
Cat. No. : RP11225
Lot. No. :PE3691805

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.103	65188	5892	1.273
2	16.006	4920236	400407	96.101
3	16.438	66350	5176	1.296
4	16.633	41595	3327	0.812
5	16.900	14752	1621	0.288
6	17.298	5568	646	0.109
7	19.393	2033	303	0.040
8	23.559	4159	465	0.081
Total		5119880	417836	100.000