


Mass Spectrum



Sample Information

User : Gary
Month-Day Processed : 12/07/16
Time Processed : 17:05:08
Injection Volume : 0.2
Sample Name : Kemptide
Cat No. : RPI0287
Lot No. : PE3531612
Theoretical MW : 771.90
Observed MW : 771.80

Interface : ESI
Nebulizing Gas Flow : 1.5L/min
CDL Temp : 250°C
Block Temp : 200°C

Interface Bias : +4.5kV
Drying Gas Flow : 5 L/min
T.Flow : 0.2ml/min
B.conc : 50%H₂O/50%MeOH