

Sample Name :Magainin 2
 Cat. No. : RP11232
 Lot. No. : PE4351902
 Time Processed : 9:25:29
 Month-Day-Year Processed : 04/20/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0309

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.462	1933	211	0.044
2	13.627	5152	289	0.117
3	14.116	3978	862	0.091
4	15.126	35706	3198	0.813
5	15.414	3161	594	0.072
6	16.148	4223568	543079	96.226
7	16.517	81147	7809	1.849
8	16.692	19082	2833	0.435
9	17.650	11030	2085	0.251
10	17.846	2096	211	0.048
11	20.704	1315	94	0.030
12	23.911	1061	164	0.024
Total		4389228	561430	100.000