

Sample Name :Exendin (9-39)
 Cat No. :RP10872
 Lot No. :E8498720E
 Time Processed : 2:28:28
 Month-Day-Year Processed : 05/20/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment:GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.881	378286	43002	2.926
2	13.427	70071	12865	0.542
3	13.617	53540	15219	0.414
4	13.812	12377988	1543242	95.740
5	14.537	48858	6174	0.378
Total		12928743	1620502	100.000