

Sample Name :InsB 9-23
 Cat. No. : RP20391
 Lot. No. : PE9261812
 Time Processed : 19:53:13
 Month-Day-Year Processed : 12/16/2018

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
32.00	Pumps	Pump B Conc.	95	
32.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.10	Controller	Stop		


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Equipment: G09010878

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.225	5900936	729445	97.892
2	11.433	118020	19516	1.958
3	11.600	9031	1131	0.150
Total		6027988	750092	100.000