

Sample Name : [Arg⁸]-Vasotocin
Cat. No. : RP10061
Lot. No. : PE2711708


Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.702	71356	10528	2.964
2	12.308	2310878	360704	95.995
3	12.590	8719	1251	0.362
4	12.867	1375	249	0.057
5	16.114	6501	660	0.270
6	16.646	2600	324	0.108
7	24.391	5862	634	0.244
Total		2407292	374349	100.000